Evaluation of Image Signal-to-Noise Ratio in Time-of-flight PET

Enrico Clementel *Student Member, IEEE*, Stefaan Vandenberghe, Joel S.Karp *Senior Member, IEEE*, Suleman Surti *Senior Member, IEEE*

Abstract—In PET imaging Noise Equivalent Counts (NEC) is a common image quality index, derived to be proportional to image SNR² and used as an index of general system performance. Many studies have shown that TOF information leads to reduced noise, faster image convergence, and improved SNR. However, the original NEC formula does not account for any contribution of timing resolution to the final image quality, and modified versions of the formula have been proposed to account for the reduction in noise variance and increased sensitivity due to TOF information. In this study, we aim to investigate the relationship between NEC and image SNR in uniform phantoms when OSEM and TOF OSEM image reconstruction is applied. Two cylindrical uniform phantoms, 20 and 35 cm in diameter, were acquired over a wide range of activity levels on a Philips Gemini TF PET scanner. Multiple realizations of the original scans were obtained through bootstrapping and reconstructed with OSEM and TOF OSEM algorithms to obtain mean and standard deviation images. The ratio of the mean value in a central ROI over both images was taken as a measure of image SNR. NEC was calculated from the original data using both the classical and a TOF-adapted formula. The results show that Trues have a better proportionality with image SNR² than NEC and TOF NEC in the considered range of activities. Timing resolution and Random fraction appear to have a limited influence on image SNR, but the usage of a matched TOF kernel in reconstruction is found to be necessary to maximize the gain.

I. INTRODUCTION

NOISE Equivalent Counts (NEC) is a common image quality index in PET, derived by Strother [1] to be proportional to the square of image Signal-to-Noise Ratio (SNR) and is used as an index of general system performance. It is given by:

$$NEC = \frac{T^2}{T + S + \frac{D}{D_{FOV}} \cdot R} \quad (1)$$

where T represents the Trues, S the Scatter and R the Random coincidences, D represents the object diameter, D_{FOV} the active Field-of-View diameter.

As clinical Time-of-Flight (TOF) PET scanners are now

available, many studies have shown that TOF information leads to reduced noise, faster image convergence, and improved SNR [2], [3], [4], [5], [6], [7], [8], [9], [10], [11]. The original NEC formula does not account for any contribution of timing resolution to the final image quality. A formula incorporating the effects of TOF information (reduction in noise variance and increased sensitivity, [12], [13]) was proposed by Conti [14]:

$$TOF NEC = \frac{D}{\Delta x} \cdot \frac{T^2}{T + S + (\frac{D}{D_{FOV}})^2 \cdot R} \quad (2)$$

where Δx represents the positioning uncertainty determined by the systems timing resolution.

At first approximation, based on formulas (1) and (2), the TOF SNR gain due to increased sensitivity is predicted to be proportional to the object dimension:

$$TOF SNR Gain \propto \frac{D}{\Delta x} \quad (3)$$

the aim of our study is to investigate the relationship between NEC and image SNR in uniform phantoms when OSEM and TOF OSEM image reconstruction is applied, and to study the influence of timing resolution and Random Fraction on image SNR.

II. MATERIALS AND METHODS

A. Data Acquisition

Two cylindrical phantoms, 30 cm long, 20 and 35 cm in diameter, were filled with a uniform solution of FDG and water. The initial activity concentrations were 6.216 kBq/ml (0.168 μ Ci/ml) for the 35 cm phantom and 22.2 kBq/ml (0.60 μ Ci/ml) for the 20 cm phantom. Multiple scans were acquired in list mode format during activity decay and final activity concentrations were 0.407 kBq/ml (0.014 μ Ci/ml) for the 35 cm phantom and 0.518 kBq/ml (0.011 μ Ci/ml) for the 20 cm phantom. Timing resolution varied between 610 and 800 ps in the considered range of activities due to degradation effects occurring at high count rates in our scanner, a Philips Gemini TF [2].

B. List mode data generation

Different list mode data were considered in this study:

Original data: The original datasets acquired on the scanner at varying count rate for each phantom.

Frac₈₀₀: A high count rate list mode dataset, for which the original timing resolution was 800 ps, was

Manuscript received November 10, 2010.

E. Clementel and S. Vandenberghe are with the Department of Electronics and Information Systems, MEDISIP, Ghent University-IBBT-IBiTech, De Pintelaan 185 block B, B-9000 Ghent, Belgium (telephone: +39-9-3324322, e-mail: enrico.clementel@ugent.be).

S. Surti is with the Department of Radiology of the University of Pennsylvania, 404 Blockley Hall, 425 Guardian Drive, Philadelphia, PA 19104 USA (telephone: 215-662-7214, e-mail: surti@mail.med.upenn.edu).

Joel S. Karp is with the Department of Physics and Astronomy and the Department of Radiology of the University of Pennsylvania, 404 Blockley Hall, 425 Guardian Drive, Philadelphia, PA 19104 USA (telephone: 215-662-3073, e-mail: joelkarp@mail.med.upenn.edu).

fractionated in 5 smaller datasets containing 12%, 26%, 38%, 57%, 78% of the original dataset's True coincidences. The data were reconstructed with TOF OSEM with a matching 800 ps TOF kernel, as well as with non-TOF OSEM. These additional datasets were generated with the purpose of comparison of a high count rate dataset with increased Random fraction and Scatter fraction, worse timing resolution but short scan time with lower activity data containing an equal amount of Trues but better timing resolution and lower Random fraction and Scatter fraction.

Frac₆₁₀: The same fractionated data from above (Frac₈₀₀) were reconstructed with the intrinsic TOF kernel of the scanner (610 ps), to investigate the effect of ignoring the degrading effects of timing resolution occurring at high count rates.

Degr₈₀₀: The timing resolution of 5 selected original datasets was degraded via software to 800 ps and reconstructed with a matching TOF kernel. Timing resolution for the original data points was 610, 640, 660, 700 and 750 ps. These data were generated for comparison with the original data, which have identical Random fraction, Scatter fraction and scan time but better timing resolution.

C. Bootstrapping

Multiple realizations of each list dataset were obtained following Dahlbom's bootstrap approach [15]. The optimal number of bootstrapped realizations was determined prior to the study: 120 bootstrapped copies of the 35 cm phantom data set were generated and reconstructed, and 12 Standard Deviation images SD were calculated using 10, 20, ..., 120 of the bootstrapped copies. The sum of all pixels in each SD image was taken as a figure of merit representing the total SD, and its convergence was taken as an indication of the minimum number of bootstrapped datasets necessary to accurately replicate the noise statistic of the original dataset. Based on this study we chose to bootstrap 60 copies per each dataset, as a good convergence of SD values is obtained for both TOF and non-TOF images.

D. Image reconstruction algorithm

An OSEM and TOF OSEM list mode algorithm with 33 subsets was used for image reconstruction. Corrections for attenuation (CT based), detector efficiency/normalization, Scatter (TOF Single Scatter Simulation [16], [17]), and Random coincidences (Delayed Coincidence Window) were incorporated in the system model. To determine the number of iterations to be used in the study, both phantoms, containing a set of six 1 cm spheres, were filled with a total activity of 111 MBq and an 8:1 sphere-to-background ratio. Based on the Contrast Recovery Coefficient curves, 3 and 6 iterations respectively for TOF and non-TOF image reconstruction were chosen for the 20 cm phantom, and 2 and 5 iterations for TOF and non-TOF were chosen for the 35 cm phantom, in order to achieve similar contrast values close to convergence. The

Fig. 1: (a) Contrast Recovery Coefficient curves for the 35 cm phantom for TOF, in red, and non-TOF reconstruction, in blue; (b) corresponding image. Four of the hot lesions are visible.

Fig. 2: Example of mean (*top*) and Standard Deviation (*bottom*) for the 35 cm phantom (*left*) and the 20 cm phantom (*right*)

CRC curve and a transaxial image of the 35 phantom used for its calculation is shown in Figure 1.

E. SNR measurement and NEC calculation

For each dataset of the original and additional data, Mean and Standard Deviation (SD) images were obtained from the 60 bootstrap realizations; an example of such images is shown in Figure 2: visual assessment shows that a good level of uniformity is achieved for both phantoms indicating accurate corrections for Random and Scatter coincidences. SNR was measured as the ratio between the mean values in a 4 cm ROI centered on both images. TOF SNR Gain was also calculated as the ratio:

$$Gain = \frac{TOF\ SNR^2}{nonTOF\ SNR^2} \quad (4)$$

III. RESULTS AND DISCUSSION

Figure 3 shows the SNR^2 vs NEC and SNR^2 vs Trues plot for the 20 cm phantom. A linear correlation is clear in both plots for the original data and for the Frac₈₀₀ data. Figure

Fig. 3: Results for the 20 cm phantom images reconstructed without TOF information; *brown squares*: original data, *orange squares*: Frac₈₀₀ data.

Top: image SNR^2 is plotted against NEC, calculated with formula 1; *bottom*: image SNR^2 is plotted against Trues.

4 shows the corresponding plots for the 35 cm phantom. SNR^2 shows a non-linear behavior against NEC. In particular, Frac₈₀₀ data, which contains a high fraction of Randoms, has the same NEC as a corresponding original dataset but higher SNR^2 . The correlation of image SNR^2 with Trues, however, is still linear. This suggests that randoms variance do not significantly affect image SNR and that the NEC formula seems to overestimate noise propagation when a blob-based OSEM reconstruction algorithm is applied, as it potentially better controls noise propagation. This is clearly observed in the larger object but is not apparent for the data of the smaller one: in our experimental setup, at the highest activity level and considering the size of the 20 cm diameter phantom, the randoms fraction is limited to 30%, and noise propagation is therefore also limited. NEC and Trues do not differ significantly and both show a good relationship with image SNR^2 . In the case of the 35 cm phantom data, however, we observe a considerably higher randoms fraction of 70% at the highest activity point. In the latter case, noise propagation could potentially have a larger effect on the final image and the differences between SNR^2 as predicted by NEC and SNR^2 as predicted by Trues are clearly visible.

Fig. 4: Results for the 35 cm phantom images reconstructed without TOF information; *brown squares*: original data, *orange squares*: Frac₈₀₀ data.

Top: image SNR^2 is plotted against NEC, calculated with formula 1; *bottom*: image SNR^2 is plotted against Trues.

Figure 5 shows the plot for TOF SNR^2 vs Trues for both phantoms, and non-TOF SNR^2 for visual reference. A good proportionality is observed for the original data for both objects. Degr₈₀₀ data line up well with the original data, suggesting timing resolution influence on image SNR is limited in the considered activity range. Frac₈₀₀ also lines up with the original data. As these data have a comparable amount of Trues but a higher Random fraction than the corresponding underlying original data point, this suggests that the Random fraction influence on image SNR is also limited in the considered activity range. Frac₆₁₀ proportionality with the Trues is accurate but with a shallower slope than the original data reconstructed with a matching TOF kernel, suggesting the use of a matched TOF kernel is beneficial in terms of image SNR. The measured TOF SNR Gain is 2.1 for the 35 cm phantom and 1.5 for the 20 cm phantom. The ratio of the two Gain measures is 1.4, which is lower than the ratio of the two diameters (1.75).

Figure 6 shows the plot for TOF SNR^2 vs TOF NEC calculated according to the formula in (2). No proportionality is observed between TOF SNR^2 and TOF NEC for the original data on both phantoms. This is more clearly visible for

Fig. 5: *Top*: SNR^2 vs Trues for the 20 cm phantom data, reconstructed with TOF OSEM; *Bottom*: SNR^2 vs Trues for the 35 cm phantom data, reconstructed with TOF OSEM;

35 cm phantom. Degr_{800} and Frac_{800} data for the 20 cm phantom follow a similar trend as the original data. Frac_{610} data for the 20 cm phantom shows a similar trend but with a shallower slope, indicating lower SNR than the corresponding original data with same TOF NEC. Frac_{800} data for the 35 cm phantom show higher image SNR^2 than the corresponding underlying original data having the same TOF NEC. The same consideration applies to Degr_{800} . Frac_{610} data follows a similar trend as the original data, has the same TOF NEC but lower SNR as the corresponding original data.

IV. CONCLUSIONS

Under our experimental conditions, Trues are found to have a better proportionality with image SNR^2 than NEC with both TOF and non-TOF OSEM image reconstruction in the considered range of activities. TOF SNR^2 gain is larger for the larger cylinder but does not increase linearly with the diameter as predicted by equation (3).

Figure 5 clearly shows the advantage in terms of image SNR when using TOF, as pointed by the steeper slope of the TOF reconstructed data line. However, as the additional data line up with the original data, the influence of timing resolution and Random fraction on image SNR seems to be limited in the ranges considered. The systematically shallower slope of the Frac_{610} data indicates that the use of matched TOF kernel during reconstruction is required to maximize the gain.

Fig. 6: *Top*: SNR^2 vs TOF NEC for the 20 cm phantom data, reconstructed with TOF OSEM; *Bottom*: SNR^2 vs TOF NEC for the 35 cm phantom data, reconstructed with TOF OSEM;

These results are focused on image pixel SNR and reflect the influence of TOF on the image pixel SNR only. We do not consider the improvements in lesion contrast and detectability that have been previously associated with TOF image reconstruction.

For the purpose of this study we choose a constant iteration number for TOF and non-TOF image reconstruction for varying count levels and count statistics. Although we expect data to require a different number of iterations to achieve contrast convergence at different count rates and count statistics, calculating the optimal number of iterations for each of the datasets would be impractical; moreover, the optimal number of iterations would depend not only on count rate, but also on the image quality index considered for convergence, as lesion detectability and contrast recovery usually converge at different iterations number. Our setup choice provides similar CRC values for TOF and non-TOF images and corresponds to the clinical approach of setting imaging reconstruction parameters to fixed values.

ACKNOWLEDGMENT

The authors would like to thank Matt Werner for his support and insight on image reconstruction, and Randy Kulp for his help with data acquisition. The authors would also like to acknowledge the NIH grants R01-CA113941 (J S Karp) and R01-EB009056 (S Surti).

REFERENCES

- [1] S. Strother, M. Casey, and E. Hoffman, "Measuring PET scanner sensitivity: relating countrates to image signal-to-noise ratios using noise equivalents counts," *Nuclear Science, IEEE Transactions on*, vol. 37, no. 2, pp. 783–788, 2002.
- [2] S. Surti, A. Kuhn, M. E. Werner, A. Perkins, J. Kolthammer, and J. S. Karp, "Performance of Philips Gemini TF PET/CT scanner with special consideration for its time-of-flight imaging capabilities," *Journal of Nuclear Medicine*, vol. 48, no. 3, pp. 471–80, march 2007.
- [3] B. Jakoby, Y. Bercier, M. Conti, M. Casey, T. Gremillion, C. Hayden, B. Bendriem, and D. Townsend, "Performance investigation of a time-of-flight PET/CT scanner," in *Nuclear Science Symposium Conference Record, 2008. NSS'08. IEEE*. IEEE, 2009, pp. 3738–3743.
- [4] B. Kemp, J. Williams, R. Ruter, V. Lowe, and B. Mullan, "Performance measurements of a whole body PET/CT system with time-of-flight capability," in *Society of Nuclear Medicine Annual Meeting Abstracts*, vol. 50, no. Supplement 2. Soc Nuclear Med, 2009, p. 1546.
- [5] S. Surti, S. Karp, L. Popescu, E. Daube-Witherspoon, and M. Werner, "Investigation of time-of-flight benefit for fully 3-DPET," *Medical Imaging, IEEE Transactions on*, vol. 25, no. 5, pp. 529–538, 2006.
- [6] M. Conti, B. Bendriem, M. Casey, M. Chen, F. Kehren, C. Michel, and V. Panin, "First experimental results of time-of-flight reconstruction on an LSO PET scanner," *Physics in Medicine and Biology*, vol. 50, p. 4507, 2005.
- [7] S. Vandenberghe, J. S. Karp, and I. Lemahieu, "Influence of TOF resolution on object dependent convergence in iterative reconstruction," *Journal of Nuclear Medicine*, vol. 47, p. 58P, 2006.
- [8] J. S. Karp, S. Surti, M. E. Daube-Witherspoon, and G. Muehllehner, "Benefit of Time-of-Flight in PET: Experimental and clinical results," *Journal of Nuclear Medicine*, vol. 49, no. 3, pp. 462–470, march 2008.
- [9] C. Lois, B. Jakoby, M. Long, K. Hubner, D. Barker, M. Casey, M. Conti, V. Panin, D. Kadrmas, and D. Townsend, "An assessment of the impact of incorporating time-of-flight information into clinical PET/CT imaging," *Journal of Nuclear Medicine*, vol. 51, no. 2, p. 237, 2010.
- [10] S. Surti and J. Karp, "Experimental evaluation of a simple lesion detection task with time-of-flight PET," *Physics in medicine and biology*, vol. 54, p. 373, 2009.
- [11] D. Kadrmas, M. Casey, M. Conti, B. Jakoby, C. Lois, and D. Townsend, "Impact of time-of-flight on PET tumor detection," *Journal of Nuclear Medicine*, vol. 50, no. 8, p. 1315, 2009.
- [12] T. Budinger, "Time-of-flight positron emission tomography: status relative to conventional PET," *Journal of nuclear medicine*, vol. 24, no. 1, p. 73, 1983.
- [13] T. Tomitani, "Image reconstruction and noise evaluation in photon time-of-flight assisted positron emission tomography," *Nuclear Science, IEEE Transactions on*, vol. 28, no. 6, pp. 4581–4589, 2007.
- [14] M. Conti, "Effects of randoms on signal-to-noise-ratio in TOF PET," *IEEE Nuclear Science Symposium Medical Imaging Conference Record*, vol. 1, pp. 1590–1595, 2005.
- [15] M. Dahlbom, "Estimation of image noise in pet using the bootstrap method," *IEEE Trans Nucl Sci*, vol. 49, pp. 2062–2066, 2002.
- [16] C. C. Watson, "Extension of Single Scatter Simulation to Scatter Correction of Time-of-flight PET," *IEEE Transactions on Nuclear Science*, vol. 54, no. 5, pp. 1679–86, October 2007.
- [17] M. E. Werner, S. Surti, and J. S. Karp, "Implementation and Evaluation of a 3D PET Single Scatter Simulation with TOF Modeling," *IEEE NSS Conference Record*, vol. 1, pp. 1768–1773, 2006.